



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/816,641	04/02/2004	Curtis G. Neason	066243-0236 (141210)	2877
33679	7590	10/11/2006	EXAMINER	
GE MEDICAL SYSTEM C/O FOLEY & LARDNER LLP 777 EAST WISCONSIN AVENUE MILWAUKEE, WI 53202-5306			WEATHERBY, ELLSWORTH	
			ART UNIT	PAPER NUMBER
			3768	

DATE MAILED: 10/11/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

**Application No.**

10/816,641

**Applicant(s)**

NEASON, CURTIS G.

**Examiner**

Ellsworth Weatherby

**Art Unit**

3768

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 14 September 2005.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-26 is/are pending in the application.
- 4a) Of the above claim(s) 18-24 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-17, 25 and 26 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)            | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)   | Paper No(s)/Mail Date. _____                                      |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>04/02/2004</u> .  | 6) <input type="checkbox"/> Other: _____                          |

## **DETAILED ACTION**

### ***Election/Restrictions***

1. Applicant's election with traverse of Group I, i.e., claims 1-17 and 25 in the reply filed on 9/14/2005 is acknowledged. The traversal is on the ground(s) that that the two grouping of claims would not be unduly burdensome to search and examine because much of the art is likely to overlap between the different groups of claims since all of the groups of claims bear some relation to an electrophysiology system. This is not found persuasive because the applicant has not addressed the requirement for restriction where, in this case, the apparatus as claimed can be used to practice another and materially different process such as acquiring the image in any lab of a hospital, including a radiology lab, and performing all the disclosed processes at the local site.

The requirement is still deemed proper and is therefore made FINAL.

### ***Claim Rejections - 35 USC § 103***

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Art Unit: 3768

3. Claims 1-3, 5-12, 14-17, 25, and 26 are rejected under 35 U.S.C. 103(a) as being unpatentable over Hu et al. (USPN 6,408,043) in view of Ben-Haim (USPN 5,391,199).

Regarding claims 1-3, 5-9, 25, and 26, Hu et al. '043 teaches volumetric computed tomography for cardiac imaging including a system for communicating the image data over any suitable network, such as the Internet (col. 10, lines 16-29).

Hu et al. '043 also discloses a data processing system configured to be communicatively coupled to one or more internal medical imaging systems and the network (figure 4, ref. 1010); A display communicatively coupled to the data processing system and configured to display a three dimensional image of the heart (cols. 4-5, lines 66-67 & 1-14); and a service facility on the network that provides remote monitoring, remote system control, immediate file access from remote locations, remote file storage and archiving, thereby allowing an image of the heart to be produced over a network (col. 9, lines 4-10).

Hu et al. '043 does not teach gathering electrical and positional information using a probe that is communicatively coupled to the data processing system.

Hu et al. '043 also does not teach storing the image data on a data storage system coupled to the network before the one or more probes are positioned inside the heart.

Hu et al. '043 also does not teach generating a report which comprises the electrical information and the position information sensed by the probe and the image acquired by the internal medical imaging system.

Ben-Haim '199 discloses an intra-cardiac probe that encompasses four aspects: the first is intended to process locating information; the second processes sensed electrical information; the third integrates previously processed information; and the fourth processes the integrated information to generate a topographical map of the sensed variable (col. 5, lines 15-21).

Ben-Haim '199 further discloses: A basic image will be recorded at the beginning of the procedure to allow determination of the cardiac chamber anatomy and of the positions of reference catheters in the heart. This image will be used thereafter as the basic source of information to describe the heart chamber anatomy (col. 15, lines 3-5). The electrophysiological signal processor will acquire electrical information from Intra-cardiac electrograms (col. 6, lines 9-11). The output device will use a computer screen or a holographic imaging unit that will be updated on a beat-by-beat base. The output will include the following information: superimposed on the basic image the position of the catheter will be represented as a symbol on the ventricular wall. The activation maps will be plotted and overlaid on the same image (col. 9, lines 32-38).

It would have been obvious to one of ordinary skill in the art at the time of applicant's invention to combine the cardiac imaging system including a system for communication of data over a network of Hu et al. '043 with the trackable probe of Ben-Haim '199. The motivation to do so would be to enhance the of the

Art Unit: 3768

invention of Hu et al. '043 to include electrical activation mapping and help study the dynamics of the chamber morphology as a function of the cardiac cycle. This simple addition to the data acquisition system of Hu et al. '043 to accept data from the intra-cardiac probe of Ben-Haim '199 facilitates increased diagnostics for more users than those at the local site, which is an obvious goal of the invention of Hu et al. '043.

Regarding claims 10-12 and 14-17, Hu et al. '043 discloses as noted above. Hu et al. '043 further discloses a processor for processing image data acquired from the internal medical imaging system where the processor is configured to be communicatively coupled over a network to a file server or some other data storage medium for storing image data files (col. 12, lines 21-42); A display that is communicatively coupled to the data processing system and configured to display an image of the heart (cols. 4-5, lines 66-67 & 1-14); A service system or file server on the network that provides remote monitoring and immediate file access from remote locations, thus allowing cardiac imaging by retrieval of the cardiac data from the file server by way of the network (col. 9, lines 4-10). Hu et al. '042 also discloses acquiring several data sets corresponding to different phases of the heart and communicating the data to a remote service facility over a network (col. 15, lines 18-23).

Hu et al. '043 does not teach gathering electrical and position information using an intra-cardiac probe that is communicatively coupled to the processor that is configured to process the electrical and position information.

Art Unit: 3768

Hu et al. '043 also does not teach acquiring imaging data from the internal medical imaging system before the one or more probes is positioned inside the heart, and wherein after the image data from the intra-cardiac probe is acquired it is stored on the file server.

Hu et al. '043 also does not teach generating a report which comprises the electrical information and the position information sensed by the probe and the image acquired by the internal medical imaging system.

Hu et al. '043 also does not disclose that the processor is used to process the position information from the probes to create a structural map of the heart.

Ben-Haim '199 discloses an intra-cardiac probe that is configured to be positioned in heart. Ben-Haim '199 also discloses a location and image processor that identifies the location of chamber boundaries using the methods of edge enhancement and edge detection, catheter locations relative to the chamber boundaries, and the dynamics of chamber morphology as a function of the cardiac cycle (col. 5, lines 41-47). The probe also acquires electrical information from intra-cardiac electrograms (col. 6, lines 9-11). Ben-Haim '199 further discloses, as noted above, acquiring and saving image data with an internal medical imaging system to establish a "basic image" before the probe is positioned inside the heart (col. 15, lines 3-5). Ben-Haim '199 also teaches generating a report that comprises electrical information, position information and the image (col. 9, lines 32-38).

It would have been obvious to one of ordinary skill in the art at the time of applicant's invention to combine the cardiac imaging system including a system

Art Unit: 3768

for communication of data over a network of Hu et al. '043 with the intra-cardiac probe of Ben-Haim '199. The motivation to do so would be to enhance the of the invention of Hu et al. '043 to include electrical activation mapping and help study the dynamics of the chamber morphology as a function of the cardiac cycle. This simple addition to the data acquisition system of Hu et al. '043 to accept data from the intra-cardiac probe of Ben-Haim '199 facilitates increased diagnostics for more users than those at the local site, which is an obvious goal of the invention of Hu et al. '043.

4. Claims 4 and 13 are rejected under 35 U.S.C. 103(a) as being unpatentable over Hu et al. '043 in view of Ben-Haim '199 as applied to claims 1 and 10 above, and further in view of Avinash et al. (PGPUB 2004/0122787).

The modified invention of Hu et al. '043 describes the invention substantially as claimed except for the network comprising a wireless network.

Avinash et al. '787 discloses an enhanced computer-assisted medical data processing system where data is linked between interfaces on any suitable network including a wireless network (0058).

It would have been obvious to one of ordinary skill in the art to combine the modified invention of Hu et al. '043 to include a wireless network as taught by Avinash et al. '787. The motivation to do so would be to facilitate data transfer through the most suitable network.




**Conclusion**

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Ellsworth Weatherby whose telephone number is (571) 272-2248. The examiner can normally be reached on M-F 8:30 a.m. - 5:00 p.m..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eleni Mantis-Mercader can be reached on (571) 272-4740. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

EW

  
BRIAN L. CASLER  
SUPERVISORY PATENT EXAMINER  
TECHNOLOGY CENTER 3700

Ellsworth Weatherby  
Examiner  
Art Unit 3768